

PATENT COOPERATION TREATY

PCT

INTERNATIONAL PRELIMINARY REPORT ON PATENTABILITY  
(Chapter I of the Patent Cooperation Treaty)

(PCT Rule 44bis)

Applicant's or agent's file reference B070580	<b>FOR FURTHER ACTION</b>		See item 4 below
International application No. PCT/JP2008/069327	International filing date ( <i>day/month/year</i> ) 24 October 2008 (24.10.2008)	Priority date ( <i>day/month/year</i> ) 27 October 2007 (27.10.2007)	
International Patent Classification (8th edition unless older edition indicated) See relevant information in Form PCT/ISA/237			
Applicant KANEKA CORPORATION			

<p>1. This international preliminary report on patentability (Chapter I) is issued by the International Bureau on behalf of the International Searching Authority under Rule 44 bis.1(a).</p> <p>2. This REPORT consists of a total of 5 sheets, including this cover sheet.</p> <p>In the attached sheets, any reference to the written opinion of the International Searching Authority should be read as a reference to the international preliminary report on patentability (Chapter I) instead.</p>																								
<p>3. This report contains indications relating to the following items:</p> <table style="width: 100%; border: none;"> <tr> <td style="width: 10%; text-align: center;"><input checked="" type="checkbox"/></td> <td style="width: 30%;">Box No. I</td> <td>Basis of the report</td> </tr> <tr> <td style="text-align: center;"><input type="checkbox"/></td> <td>Box No. II</td> <td>Priority</td> </tr> <tr> <td style="text-align: center;"><input type="checkbox"/></td> <td>Box No. III</td> <td>Non-establishment of opinion with regard to novelty, inventive step and industrial applicability</td> </tr> <tr> <td style="text-align: center;"><input checked="" type="checkbox"/></td> <td>Box No. IV</td> <td>Lack of unity of invention</td> </tr> <tr> <td style="text-align: center;"><input checked="" type="checkbox"/></td> <td>Box No. V</td> <td>Reasoned statement under Article 35(2) with regard to novelty, inventive step or industrial applicability; citations and explanations supporting such statement</td> </tr> <tr> <td style="text-align: center;"><input type="checkbox"/></td> <td>Box No. VI</td> <td>Certain documents cited</td> </tr> <tr> <td style="text-align: center;"><input type="checkbox"/></td> <td>Box No. VII</td> <td>Certain defects in the international application</td> </tr> <tr> <td style="text-align: center;"><input type="checkbox"/></td> <td>Box No. VIII</td> <td>Certain observations on the international application</td> </tr> </table> <p>4. The International Bureau will communicate this report to designated Offices in accordance with Rules 44bis.3(c) and 93bis.1 but not, except where the applicant makes an express request under Article 23(2), before the expiration of 30 months from the priority date (Rule 44bis .2).</p>	<input checked="" type="checkbox"/>	Box No. I	Basis of the report	<input type="checkbox"/>	Box No. II	Priority	<input type="checkbox"/>	Box No. III	Non-establishment of opinion with regard to novelty, inventive step and industrial applicability	<input checked="" type="checkbox"/>	Box No. IV	Lack of unity of invention	<input checked="" type="checkbox"/>	Box No. V	Reasoned statement under Article 35(2) with regard to novelty, inventive step or industrial applicability; citations and explanations supporting such statement	<input type="checkbox"/>	Box No. VI	Certain documents cited	<input type="checkbox"/>	Box No. VII	Certain defects in the international application	<input type="checkbox"/>	Box No. VIII	Certain observations on the international application
<input checked="" type="checkbox"/>	Box No. I	Basis of the report																						
<input type="checkbox"/>	Box No. II	Priority																						
<input type="checkbox"/>	Box No. III	Non-establishment of opinion with regard to novelty, inventive step and industrial applicability																						
<input checked="" type="checkbox"/>	Box No. IV	Lack of unity of invention																						
<input checked="" type="checkbox"/>	Box No. V	Reasoned statement under Article 35(2) with regard to novelty, inventive step or industrial applicability; citations and explanations supporting such statement																						
<input type="checkbox"/>	Box No. VI	Certain documents cited																						
<input type="checkbox"/>	Box No. VII	Certain defects in the international application																						
<input type="checkbox"/>	Box No. VIII	Certain observations on the international application																						

<p>The International Bureau of WIPO 34, chemin des Colombettes 1211 Geneva 20, Switzerland</p> <p>Facsimile No. +41 22 338 82 70</p>	<table border="1" style="width: 100%;"> <tr> <td>Date of issuance of this report 27 April 2010 (27.04.2010)</td> </tr> <tr> <td>Authorized officer  <b>Yoshiko Kuwahara</b></td> </tr> <tr> <td>e-mail: pt07.pct@wipo.int</td> </tr> </table>	Date of issuance of this report 27 April 2010 (27.04.2010)	Authorized officer  <b>Yoshiko Kuwahara</b>	e-mail: pt07.pct@wipo.int
Date of issuance of this report 27 April 2010 (27.04.2010)				
Authorized officer  <b>Yoshiko Kuwahara</b>				
e-mail: pt07.pct@wipo.int				

# 特許協力条約

発信人 日本国特許庁（国際調査機関）

代理人 特許業務法人原謙三国際特許事務所  様  あて名 〒530-0041 日本国大阪府大阪市北区天神橋2丁目北2番6号 大和南森町ビル	PCT 国際調査機関の見解書 (法施行規則第40条の2) [PCT規則43の2.1]
発送日 (日.月.年) 02.12.2008	

出願人又は代理人 の書類記号 B070580	今後の手続きについては、下記2を参照すること。
---------------------------	-------------------------

国際出願番号 PCT/JP2008/069327	国際出願日 (日.月.年) 24.10.2008	優先日 (日.月.年) 27.10.2007
-----------------------------	-----------------------------	---------------------------

国際特許分類 (IPC) Int.Cl. A61M25/01(2006.01)i, A61M25/00(2006.01)i

出願人 (氏名又は名称)  
 株式会社カネカ

1. この見解書は次の内容を含む。

- 第I欄 見解の基礎
- 第II欄 優先権
- 第III欄 新規性、進歩性又は産業上の利用可能性についての見解の不作成
- 第IV欄 発明の単一性の欠如
- 第V欄 PCT規則43の2.1(a)(i)に規定する新規性、進歩性又は産業上の利用可能性についての見解、それを裏付けるための文献及び説明
- 第VI欄 ある種の引用文献
- 第VII欄 国際出願の不備
- 第VIII欄 国際出願に対する意見

2. 今後の手続き  
 国際予備審査の請求がされた場合は、出願人がこの国際調査機関とは異なる国際予備審査機関を選択し、かつ、その国際予備審査機関がPCT規則66.1の2(b)の規定に基づいて国際調査機関の見解書を国際予備審査機関の見解書とみなさない旨を国際事務局に通知していた場合を除いて、この見解書は国際予備審査機関の最初の見解書とみなされる。

この見解書が上記のように国際予備審査機関の見解書とみなされる場合、様式PCT/ISA/220を送付した日から3月又は優先日から22月のうちいずれか遅く満了する期限が経過するまでに、出願人は国際予備審査機関に、適当な場合は補正書とともに、答弁書を提出することができる。

さらなる選択肢は、様式PCT/ISA/220を参照すること。

3. さらなる詳細は、様式PCT/ISA/220の備考を参照すること。

見解書を作成した日  
 18.11.2008

名称及びあて先 日本国特許庁 (ISA/JP) 郵便番号100-8915 東京都千代田区霞が関三丁目4番3号	特許庁審査官 (権限のある職員) 田中 玲子 電話番号 03-3581-1101 内線 3344	3E 3620
---	--	---------

## 第 I 欄 見解の基礎

1. 言語に関し、この見解書は以下のものに基づき作成した。

出願時の言語による国際出願

出願時の言語から国際調査のための言語である \_\_\_\_\_ 語に翻訳された、この国際出願の翻訳文  
(PCT規則12.3(a)及び23.1(b))

2.  この見解書は、PCT規則 91 の規定により国際調査機関が認めた又は国際調査機関に通知された明らかな誤りの訂正を考慮して作成した (PCT規則 43 の 2.1(b))。

3. この国際出願で開示されたヌクレオチド又はアミノ酸配列に関して、以下に基づき見解書を作成した。

a. タイプ  配列表

配列表に関連するテーブル

b. フォーマット  紙形式

電子形式

c. 提出時期  出願時の国際出願に含まれていたもの

この国際出願と共に電子形式により提出されたもの

出願後に、調査のために、この国際調査機関に提出されたもの

4.  さらに、配列表又は配列表に関連するテーブルを提出した場合に、出願後に提出した配列若しくは追加して提出した配列が出願時に提出した配列と同一である旨、又は、出願時の開示を超える事項を含まない旨の陳述書の提出があった。

5. 補足意見：

## 第IV欄 発明の単一性の欠如

1.  追加手数料の納付命令書（様式PCT/ISA/206）に対して、出願人は、規定期間内に、
- 追加手数料を納付した。
  - 追加手数料及び、該当する場合には、異議申立手数料の納付と共に、異議を申し立てた。
  - 追加手数料の納付と共に異議を申し立てたが、規定の異議申立手数料を支払わなかった。
  - 追加手数料を納付しなかった。
2.  国際調査機関は、発明の単一性の要件を満たしていないと判断したが、追加手数料の納付を出願人に求めないこととした。
3. 国際調査機関は、PCT規則 13.1、13.2 及び 13.3 に規定する発明の単一性を次のように判断する。
- 満足する。
  - 以下の理由により満足しない。

請求の範囲 1-9 に係る発明の共通の事項は、バルーンと比較してカテーテル長軸方向に伸縮しないチューブが、拘束部材の一侧に偏心して設けられている点である。

しかしながら、調査の結果、その点は、JP 4-8337 A（オリンパス光学工業株式会社）1992.01.13, 第4頁左下欄第14-15行, 第4頁左下欄第19行-第5頁左上欄第11行, 第7,8図に開示されているから、新規でないことが明らかとなった。結果として、バルーンと比較してカテーテル長軸方向に伸縮しないチューブが、拘束部材の一侧に偏心して設けられている点は、先行技術の域を出ないから、PCT規則 13.2 の第2文の意味において、この共通事項は特別な技術的特徴ではない。

それ故、請求の範囲 1-9 に係る発明全てに、共通の事項は存在しない。

4. したがって、国際出願の次の部分について、この見解書を作成した。

すべての部分

請求の範囲 \_\_\_\_\_ に関する部分

第V欄 新規性、進歩性又は産業上の利用可能性についてのPCT規則43の2.1(a)(i)に定める見解、それを裏付ける文献及び説明

1. 見解

新規性 (N)	請求の範囲	5, 9	有
	請求の範囲	1-4, 6-8	無
進歩性 (I S)	請求の範囲	5	有
	請求の範囲	1-4, 6-9	無
産業上の利用可能性 (I A)	請求の範囲	1-9	有
	請求の範囲		無

2. 文献及び説明

文献1 : JP 4-8337 A (オリンパス光学工業株式会社) 1992. 01. 13, 第4頁左下欄第14-15行, 第4頁左下欄第19行-第5頁左上欄第11行, 第7, 8図

請求の範囲1-4, 6-8に係る発明は、国際調査で引用された上記文献1 (第4頁左下欄第14-15行, 第4頁左下欄第19行-第5頁左上欄第11行, 第7, 8図) により新規性・進歩性を有しない。文献1には、処置具挿通チャンネル37が、コイルスプリング43と保護チューブ34の一侧に偏心して設けられている点が記載されている。ここで、保護チューブ34は処置具挿通チャンネル37の偏心する側と近接していることからみて、保護チューブ34は処置具挿通チャンネル37の偏心する側に設けられているといえる。

請求の範囲5に係る発明は、国際調査報告に引用されたいずれの文献にも記載されておらず、当業者にとって自明なものでもない。

請求の範囲9に係る発明は、文献1より進歩性を有しない。処置具挿通チャンネル37の強度を高めるべく補強層を備えることは、当業者が適宜選択し得る設計事項に過ぎない。